

**SERTIFIKAT HASIL PENGUJIAN**  
**No. SHP/045/VI/2024**

Product : Timbangan  
 Brand : OHAUS  
 Type : **NV2201**  
 Nomor Seri : 8345052748  
 Kapasitas : 2.200 g  
 Ketelitian : 0.1 g

**CALIBRATION**

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	2000	2000,1	0.1	±0.2	<b>BAIK</b>

**VERIFICATION**

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	100	100,0	0.0	±0.2	<b>BAIK</b>
2	200	200,0	0.0	±0.2	
3	500	500,1	0.1	±0.2	
4	1000	1000,1	0.1	±0.2	
5	2000	2000,0	0.0	±0.2	

**Catatan :**

- Standart yang digunakan : Class F1 Tolerance
- Metode Pengujian : Perbandingan langsung (Direct Comparison)
- Tanggal Pengujian : 13 Juni 2024

Tanggal diterbitkan : 14 Juni 2024  
 PT. Presisi Labtools Indonesia



**Ilham**

*Approved e-Signatory*