

SERTIFIKAT HASIL PENGUJIAN
No. SHP/046/VI/2024

Product : Timbangan
 Brand : OHAUS
 Type : **NV2201**
 Nomor Seri : 8345052746
 Kapasitas : 2.200 g
 Ketelitian : 0.1 g

CALIBRATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	2000	2000,0	0.0	±0.2	BAIK

VERIFICATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	100	100,0	0.0	±0.2	BAIK
2	200	200,0	0.0	±0.2	
3	500	500,0	0.0	±0.2	
4	1000	1000,0	0.0	±0.2	
5	2000	2000,0	0.0	±0.2	

Catatan :

- Standart yang digunakan : Class F1 Tolerance
- Metode Pengujian : Perbandingan langsung (Direct Comparison)
- Tanggal Pengujian : 13 Juni 2024

Tanggal diterbitkan : 14 Juni 2024
 PT. Presisi Labtools Indonesia



Ilham
Approved e-Signatory