

SERTIFIKAT HASIL PENGUJIAN
No. SHP/029/I/2024

Product : Moisture Analyzer
 Brand : OHAUS
 Type : **MB120**
 Nomor Seri : C334694589
 Kapasitas : 120 g
 Ketelitian : 0.001 g

CALIBRATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	50	50,000	0.000	±0.002	BAIK

VERIFICATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	50	50,000	0.000	±0.002	BAIK
2	100	100,000	0.000	±0.002	

TEMPERATURE

NO	STANDARD °C	PENUNJUKAN °C	KOREKSI °C	TOLERANSI °C	KET
1	100°C	100.0°C	0.0	±0.2	BAIK
2	160°C	160.2°C	0.2	±0.2	

Catatan :

- Standart yang digunakan : Class F1 Tolerance & Temperature Calibration
- Metode Pengujian : Perbandingan langsung (Direct Comparison)
- Tanggal Pengujian : 10 Januari 2024

Tanggal diterbitkan : 10 Januari 2024
 PT. Presisi Labtools Indonesia



Reza M
Approved e-Signatory