

SERTIFIKAT HASIL PENGUJIAN
No. SHP/042/VII/2024

Product : Timbangan
 Brand : OHAUS
 Type : **NV1202**
 Nomor Seri : 8345101104
 Kapasitas : 1.200 g
 Ketelitian : 0.01 g

CALIBRATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	1000	1000,00	0.00	±0.02	BAIK

VERIFICATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	50	50,00	0.00	±0.02	BAIK
2	100	100,00	0.00	±0.02	
3	200	200,01	0.01	±0.02	
4	500	500,01	0.01	±0.02	
5	1000	1000,00	0.00	±0.02	

Catatan :

- Standart yang digunakan : Class F1 Tolerance
- Metode Pengujian : Perbandingan langsung (Direct Comparison)
- Tanggal Pengujian : 06 Juli 2024

Tanggal diterbitkan : 18 Juli 2024
 PT. Presisi Labtools Indonesia



Reza M

Approved e-Signatory