

SERTIFIKAT HASIL PENGUJIAN
No. SHP/043/VII/2024

Product : Timbangan
 Brand : OHAUS
 Type : **PR224**
 Nomor Seri : C345102626
 Kapasitas : 200 g
 Ketelitian : 0.0001 g

CALIBRATION

| NO | STANDARD (g) | PENUNJUKAN (g) | KOREKSI (g) | TOLERANSI (g) | KET |
|----|--------------|----------------|-------------|---------------|-------------|
| 1 | 200 | 200,0000 | 0.0000 | ±0.0002 | BAIK |

VERIFICATION

| NO | STANDARD (g) | PENUNJUKAN (g) | KOREKSI (g) | TOLERANSI (g) | KET |
|----|--------------|----------------|-------------|---------------|-------------|
| 1 | 10 | 10.0002 | 0.0002 | ±0.0002 | BAIK |
| 2 | 20 | 20.0000 | 0.0000 | ±0.0002 | |
| 3 | 50 | 50.0001 | 0.0001 | ±0.0002 | |
| 4 | 100 | 100.0000 | 0.0000 | ±0.0002 | |
| 5 | 200 | 200.0000 | 0.0000 | ±0.0002 | |

Catatan :

- Standart yang digunakan : Class F1 Tolerance
- Metode Pengujian : Perbandingan langsung (Direct Comparison)
- Tanggal Pengujian : 09 Juli 2024

Tanggal diterbitkan : 18 Juli 2024
 PT. Presisi Labtools Indonesia



Reza M

Approved e-Signatory