



SURAT KETERANGAN HASIL PENGUJIAN

No. SKHP/0054/VIII/2024

Product : **TIMBANGAN**
Brand : OHAUS
Type : NVT10201
Nomor Seri : 8345274081
Capacity : 10200 g
Readability : 0.1 g

ADJUSTMENT

NO	STANDARD (g)	PENUNJUKAN (g)	DEVIASI (g)	TOLERANSI (g)	KETERANGAN
1	10000,0	10000,0	0,0	± 0,1 g	Adjustment Done
Result			Instrument Baik		

VERIFICATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KETERANGAN
1	500,0	500,1	± 0,1	± 0,1 g	Instrumen dalam kondisi baik
2	1000,0	1000,0	± 0,0	± 0,1 g	
3	2000,0	2000,1	± 0,1	± 0,1 g	
4	5000,0	5000,0	± 0,0	± 0,1 g	
5	10000,0	10000,0	± 0,0	± 0,1 g	
Result			Instrument Baik		

Catatan :

1. Standar yang digunakan : Weight Set F1 Class
2. Metode Pengujian : Pengujian Langsung
3. Tanggal Pengujian : 23 Agustus 2024

Jakarta, 27 Agustus 2024

Examiner



Muhammad Reza
Product Support