



## **SURAT KETERANGAN HASIL PENGUJIAN**

No. SKHP/0055/VIII/2024

Product : **TIMBANGAN**  
Brand : OHAUS  
Type : NVT10201  
Nomor Seri : 8345274082  
Capacity : 10200 g  
Readability : 0.1 g

### **ADJUSTMENT**

NO	STANDARD (g)	PENUNJUKAN (g)	DEVIASI (g)	TOLERANSI (g)	KETERANGAN
1	10000,0	10000,0	0,0	± 0,1 g	Adjustment Done
Result			Instrument Baik		

### **VERIFICATION**

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KETERANGAN
1	500,0	500,1	± 0,1	± 0,1 g	Instrumen dalam kondisi baik
2	1000,0	1000,0	± 0,0	± 0,1 g	
3	2000,0	2000,0	± 0,0	± 0,1 g	
4	5000,0	5000,1	± 0,1	± 0,1 g	
5	10000,0	10000,0	± 0,0	± 0,1 g	
Result			Instrument Baik		

#### **Catatan :**

- Standar yang digunakan : Weight Set F1 Class
- Metode Pengujian : Pengujian Langsung
- Tanggal Pengujian : 23 Agustus 2024

Jakarta, 27 Agustus 2024

Examiner



**Muhammad Reza**  
Product Support