

SERTIFIKAT HASIL PENGUJIAN
No. SHP/058/XII/2024

Product : Moisture Analyzer
 Brand : OHAUS
 Type : **MB95**
 Nomor Seri : C438774108
 Kapasitas : 90 g
 Ketelitian : 0.001 g

CALIBRATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	50	50,001	0.001	±0.002	BAIK

VERIFICATION

NO	STANDARD (g)	PENUNJUKAN (g)	KOREKSI (g)	TOLERANSI (g)	KET
1	50	50,000	0.000	±0.002	BAIK
2	70	70,000	0.000	±0.002	

TEMPERATURE

NO	STANDARD °C	PENUNJUKAN °C	KOREKSI °C	TOLERANSI °C	KET
1	100°C	100.1°C	0.1	±0.2	BAIK
2	160°C	160.1°C	0.1	±0.2	

Catatan :

- Standart yang digunakan : Class F1 Tolerance & Temperature Calibration
- Metode Pengujian : Perbandingan langsung (Direct Comparison)
- Tanggal Pengujian : 16 Desember 2024

Tanggal diterbitkan : 16 Desember 2024
 PT. Presisi Labtools Indonesia



M. Syahrir
Approved e-Signatory